




XactGauge™ Convection-Enhanced Pirani Gauge Sensor XGS-321



Operating Manual
Incl. EU Declaration of Conformity

Product Identification

In all communications with Ideal Vacuum Products, please specify the information given on the product nameplate. For convenient reference copy that information into the space provided below.

Ideal Vacuum Products, LCC		
PN:	
SN:	
..... VDC; W 	

Validity

This document applies to products with part numbers:

P1010261	(ISO-KF-16)
P1010260	(ISO-KF-25)
P1010259	(ISO-KF-40)
P1010258	(CF 1.33")
P1010257	(CF 2.75")
P1010256	(4 VCR female)
P1010255	(8 VCR female)
P1010245	(1/8" MNPT)

The part number (PN) can be taken from the product nameplate.

If not indicated otherwise in the legends, the illustrations in this document correspond to the product with vacuum connection DN 25 ISO-KF. They apply to the other products by analogy.

We reserve the right to make technical changes without prior notice.

Important User Information


There are operational characteristic differences between solid state equipment and electromechanical equipment. Because of these differences, and because there are a variety of uses for solid state equipment, all persons that apply this equipment must take every precaution and satisfy themselves that the intended application of this equipment is safe and used in an acceptable manner.

In no event will Ideal Vacuum Products be responsible or liable for indirect or consequential damages that result from the use or application of this equipment.

Any examples or diagrams included in this manual are provided solely for illustrative purposes. Because of the many variables and requirements imposed on any particular installation, Ideal Vacuum Products cannot assume responsibility or liability for any actual use based on the examples and diagrams.

No patent liability is assumed by Ideal Vacuum Products with respect to use of information circuits, equipment, or software described in this manual.

Throughout this manual we use notes, notices and apply internationally recognized symbols and safety messages to make you aware of safety considerations.



Identifies information about practices or circumstances that can cause electrical or physical hazards which, if precautions are not taken, could result in death or serious injury, property damage, or economic loss.



CAUTION

Identifies information about practices or circumstances that can cause electrical or physical hazards which, if precautions are not taken, could result in minor or moderate injury, property damage, or economic loss.

NOTICE

Identifies information that is critical for successful application and understanding of the product.



SHOCK HAZARD

Labels may be located on or inside the device to alert people that dangerous voltages may be present.

General Safety Instructions

- Adhere to the applicable regulations and take the necessary precautions for the process media used.
Consider possible reactions with the product materials.
Consider possible reactions (e.g. explosion) of the process media due to the heat generated by the product.
- Adhere to the applicable regulations and take the necessary precautions for all work you are going to do and consider the safety instructions in this document.
- Before beginning to work, find out whether any vacuum components are contaminated. Adhere to the relevant regulations and take the necessary precautions when handling contaminated parts.

Communicate the safety instructions to all other users.

Liability and Warranty

Ideal Vacuum Products assumes no liability and the warranty becomes null and void if the end-user or third parties


- disregard the information in this document
- use the product in a non-conforming manner
- make any kind of interventions (modifications, alterations etc.) on the product
- use the product with accessories not listed in the product documentation.

The end-user assumes the responsibility in conjunction with the process media used.

Gauge failures due to contamination or wear and tear, as well as expendable parts (e.g. Pirani filament), are not covered by the warranty.

Contents

Product Identification	2
Validity	2
Important User Information	3
General Safety Instructions	5
Liability and Warranty	5
1 Introduction / General Information	7
1.1 Description	7
1.2 Specifications	8
1.3 Dimensions	9
2 Important Safety Information	10
2.1 Safety Precautions - General	10
2.2 Safety Precautions - Service and Operation	11
2.3 Electrical Conditions	12
2.3.1 Proper Equipment Grounding	12
2.3.2 Electrical Interface and Control	13
2.4 Overpressure and use with hazardous gases	13
2.5 Gases other than Nitrogen / air	14
3 Installation	15
3.1 Mechanical Installation	15
3.2 Electrical Installation	17
3.2.1 Grounding	17
4 Using the gauge with different gases	19
5 Service	23
5.1 Calibration	23
5.2 Maintenance	23
5.3 Troubleshooting	24
5.4 Contamination	26
5.5 Gauge replacement	28
6 Factory Service and Support	29
7 Returning the Product	29
8 Disposal	30
EU Declaration of Conformity	31

For cross-references within this document, the symbol (→  XY) is used,

1 Introduction / General Information

1.1 Description


Thermal conductivity gauges measure pressure indirectly by sensing the loss of heat from a sensor to the surrounding gases. The higher the pressure of the surrounding gas, the more heat is conducted away from the sensor. Pirani thermal conductivity gauges maintain a sensor (usually a wire) at some constant temperature, and measure the current or power required to maintain that temperature. A standard Pirani gauge has a useful measuring range of about 10^{-4} Torr to 10 Torr. By taking advantage of convection currents that are generated above 1 Torr, convection-enhanced Pirani gauges increase the measuring range to just above atmosphere.

The Ideal Vacuum Products XactGauge™ XGS-321 Convection-Enhanced Pirani Gauge Sensor is a convection-enhanced Pirani vacuum gauge sensor that interfaces with external controllers such as the Ideal Vacuum Products XactGauge™ XGC-320 controller. The XGS-321 is also a direct drop-in replacement pin-pin compatible gauge for the MKS Instruments / Granville-Phillips® 275 Convectron® gauge. The sensor connector has the same pinouts and signal as the corresponding Convectron®. It is directly interchangeable with your existing Convectron® controllers and cables, so you don't need to change any wiring, hardware, or process recipes.

1.2 Specifications

Measurement range	$1.3 \times 10^{-4} \dots 1333$ mbar $1 \times 10^{-4} \dots 1000$ Torr 1.3×10^{-2} Pa ... 133 kPa
Accuracy - N ₂ (typical)	
$1.3 \times 10^{-4} \dots 1.3 \times 10^{-3}$ mbar	0.1 $\times 10^{-3}$ mbar resolution
$1.3 \times 10^{-3} \dots 530$ mbar	±10% of reading
530 ... 1333 mbar	±2.5% of reading
$1 \times 10^{-4} \dots 1 \times 10^{-3}$ Torr	0.1 mTorr resolution
$1 \times 10^{-3} \dots 400$ Torr	±10% of reading
400 ... 1000 Torr	±2.5% of reading
Repeatability - N ₂ (typical)	±2% of reading
Materials exposed to vacuum	gold-plated tungsten, 304 & 316 stainless steel, glass, nickel, Teflon®
Internal volume	26 cm ³ (1.589 in ³)
Internal surface area	59.7 cm ² (9.25 in ²)
Weight	3 oz. (85 g)
Permissible temperature	
Operating	0 ... +50 °C
Storage	−40 ... +70 °C
Bakeout temperature	≤150 °C (non-operating, with electronics cable detached)
Relative humidity	0 ... 95%, non-condensing
Use	
Operating	altitude up to 2500 m (8200 ft.)
Storage	altitude up to 12500 m (41000 ft.)
Mounting orientation	horizontal recommended (orientation has no effect on measurements below 1.3 mbar (1 Torr))

1.3 Dimensions



Fitting	Dimension A	
	mm	(inch)
DN 16 ISO-KF	33	(1.3)
DN 25 ISO-KF	33	(1.3)
DN 40 ISO-KF	33	(1.3)
DN 16 CF 1.33	27.4	(1.08)
DN 40 CF 2.75"	37.3	(1.47)
4 VCR female	47.2	(1.86)
8 VCR female	44.5	(1.75)
1/8" NPT male	25.4	(1)

2 Important Safety Information

Ideal Vacuum Products has designed and tested this product to provide safe and reliable service, provided it is installed and operated within the strict safety guidelines provided in this manual. **Please read and follow all warnings and instructions.**



WARNING

To avoid serious injury or death, follow the safety information in this document. Failure to comply with these safety procedures could result in serious bodily harm, including death, and or property damage.

Failure to comply with these warnings violates the safety standards of installation and intended use of this instrument. Ideal Vacuum Products disclaims all liability for the customer's failure to comply with these instructions.

Although every attempt has been made to consider most possible installations, Ideal Vacuum Products cannot anticipate every contingency that arises from various installations, operation, or maintenance of the module. If you have any questions about the safe installation and use of this product, please contact Ideal Vacuum Products.

2.1 Safety Precautions - General



WARNING! There are no operator serviceable parts or adjustments inside the gauge sensor.

Do not modify this product or substitute any parts without authorization of qualified Ideal Vacuum Products service trained personnel. Return the product to an Ideal Vacuum Products qualified service and repair center to ensure that all safety features are maintained. Do not use this product if unauthorized modifications have been made.



WARNING! Source power must be removed from the product prior to performing any servicing.

Prior to installing the XGS-321, ensure that all safety checks are made by a qualified service person. When a replacement gauge is required, ensure that the gauge is specified by Ideal Vacuum Products. Substitutions of non-qualified parts may result in fire, electric shock or other hazards. Use of unauthorized parts or modifications made to this product will void the warranty.

To reduce the risk of fire or electric shock, do not expose this product to rain or moisture. These products are not waterproof and careful attention must be paid to not spill any type of liquid onto these products. Do not use these products if they have been damaged. Immediately contact Ideal Vacuum Products to arrange return of the product if it is damaged.

Due to the possibility of corrosion when used in certain environmental conditions, it is possible that the product's safety could be compromised over time. It is important that the product be periodically inspected for sound electrical connections and equipment grounding. Do not use if the equipment grounding or electrical insulation has been compromised.

2.2 Safety Precautions - Service and Operation

Ensure that the vacuum port on which the XGS-321 vacuum gauge is mounted is electrically grounded.

Remove cable to the unit before attempting to service the gauge.

Remove cable to the unit if a cable or plug is damaged or the product is not operating normally according to this Operating Manual. Contact qualified Ideal Vacuum Products service personnel for any service or troubleshooting condition that may not be covered by this Operating Manual.

It is important that the product be periodically inspected for sound electrical connections and equipment grounding. Do not use if the equipment grounding or electrical insulation has been compromised.

Do not use if the unit has been dropped. Contact Ideal Vacuum Products for further instructions regarding evaluation of the damaged sensor.

2.3 Electrical Conditions



WARNING! When high voltage is present in any vacuum system, a life threatening electrical shock hazard may exist unless all exposed electrical conductors are maintained at earth ground potential. This applies to all products that come in contact with the gas contained in vacuum chambers. An electrical discharge within a gaseous environment may couple dangerous high voltage directly to any ungrounded conductor of electricity. A person could be seriously injured or killed by coming in contact with an exposed, ungrounded electrical conductor at high voltage potential. This condition applies to all products that may come in contact with the gas inside the vacuum chamber (vacuum / pressure containment vessel).

2.3.1 Proper Equipment Grounding



WARNING! Hazardous voltages that could seriously injure or cause death are present in many vacuum processes. Verify that the vacuum port on which the XGS-321 vacuum gauge module is mounted is electrically grounded. Consult a qualified Electrician if you are in doubt about your equipment grounding. Proper grounding of your equipment is essential for safety as well as intended operation of the equipment. The XGS-321 module vacuum gauge must be connected directly to a good quality earth ground. Use a ground lug on the XGS-321 gauge vacuum connection / flange if necessary.



WARNING! In order to protect personnel from electric shock and bodily harm, shield all conductors which are subject to potential high voltage electrical discharges in or around the vacuum system.

2.3.2 Electrical Interface and Control

It is the user's responsibility to ensure that the electrical signals from this product and any connections made to external devices, for example, relays and solenoids, are used in a safe manner. Always double check the system set-up before using any signals to automate your process. Perform a hazardous operation analysis of your system design and ensure safeguards and personnel safety measures are taken to prevent injury and property damage.

2.4 Overpressure and use with hazardous gases



WARNING! Install suitable protective devices that will limit the level of pressure inside your vacuum chamber to less than what the vacuum chamber system components are capable of withstanding. Ideal Vacuum Products gauges should not be used at pressures exceeding 1000 Torr absolute pressure.

In cases where an equipment failure could cause a hazardous condition, always implement fail-safe system operation. For example, use a pressure relief device in an automatic backfill operation where a malfunction could result in high internal pressures if the pressure relief device was not installed on the chamber.

The XGS-321 vacuum gauge module is not intended for use at pressures above 20 psia (1000 Torr); DO NOT exceed 35 psig (<2½ bars) pressure inside the sensor. If your chamber goes to higher pressures, you should install an isolation valve or pressure relief device to protect the gauge tube from overpressure conditions. With some fittings, actual safe overpressure conditions may be lower; for example, a quick-connect, O-ring compression fitting may forcibly release the gauge tube from the vacuum chamber fitting with only a few psi over local uncorrected barometric (atmospheric) pressure.



CAUTION! If the internal pressure of a vacuum gauge device is allowed to increase above local uncorrected barometric pressure (atmospheric pressure side), vacuum fittings may release and possible overpressure conditions may cause leaks that would allow the gas inside the gauge tube to release into the atmosphere of the surrounding environment. Toxic, pyrophoric and flammable gases are examples of hazardous gases that if allowed to leak out of the vacuum/pressure containment vessel into the atmospheric environment, could cause bodily injury and possible damage to equipment. Never expose the gauge tube internal volume to pressure above local atmospheric pressure when using hazardous gases.

2.5 Gases other than Nitrogen / air



WARNING! Do not attempt to use with gases other than nitrogen (N₂) or air without referring to correction factor data tables.

Ideal Vacuum Products gauges are calibrated for direct readout of nitrogen or air. Do not attempt to use with other gases such as argon (Ar) or carbon dioxide (CO₂) unless accurate conversion data for N₂ to other gas is properly used. Refer to the correction factor data listed in the controller Operating Manual operating this device. The Ideal Vacuum Products XGC-320 controller Operating Manuals provides a more complete discussion of using correction factors when using the gauge on gases other than Nitrogen.



WARNING! Do not use this device in an explosive atmosphere or in the presence of flammable gases, vapors or fumes. Do not use this device to measure the pressure of explosive or combustible gases or gas mixtures. The sensor wire in the gauge normally operates at 125 °C, but if malfunction should occur, the wire temperature could exceed the ignition temperature of certain combustible gases and gas mixture. This could cause an explosion which could result in serious injury or death.

3 Installation


3.1 Mechanical Installation

Mount the XGS-321 as close as possible to the pressure you want to measure. Long or restricted, small diameter tubing will create a pressure difference between your process chamber and the gauge. This may cause a delay in response to pressure changes.


Mounting the XGS-321 too close to a gas source inlet may also cause measurement and control instability. Do not mount the XGS-321 near a source of heating or cooling, such as heaters or air conditioning vents.

Mount the XGS-321 with its main (long) axis horizontal (see diagram below). Pressure reading errors may occur above 1 Torr if the unit is not mounted horizontally. Below 1 Torr, mounting position has little to no effect.

For accurate measurements above 1 Torr, mount the gauge axis horizontally as shown below:



Incorrect Orientation:



Mount the XGS-321 with port down, if possible, to help minimize the effect of any particles or condensation from collecting in the gauge.

Do not mount the XGS-321 where it will be subjected to excessive vibration. Vibrations may cause unstable readings, measurement errors and possible mechanical stress to components in the XGS-321.

Flanges/ Fittings - follow the manufacturer's recommendations and note the following:


- NPT fittings: When connecting the device using a NPT fitting, apply a thread sealant compound or wrap the threaded portion of the tubing with one-and-a-half to two wraps of pipe thread seal tape such as PTFE (Teflon®) tape and hand tighten the gauge into the gauge port. Do not use a wrench or other tool which may damage the gauge.

3.2 Electrical Installation

3.2.1 Grounding



Be sure the vacuum gauge and the rest of your vacuum system are properly grounded for safety as well as intended operation of the equipment. When using KF flanges, metal clamps must be used to ensure proper grounding. Be aware that some vacuum fittings such as NPT connections installed using Teflon tape may not allow for metal-to-metal contact between the vacuum gauge and the vacuum chamber. If such is the case, use a 12 gauge or larger copper wire to connect the vacuum gauge to a ground lug on your vacuum chamber as shown below.



4 Using the gauge with different gases

A thermal conductivity gauge senses heat loss which depends on the thermal conductivity of the gas surrounding the sensor. Since different gases, and mixtures, have different thermal conductivities, the indicated pressure readings and outputs will also be different. Ideal Vacuum Products convection gauges (and most other thermal conductivity gauges) are calibrated using nitrogen (N_2). When a gas other than N_2 / air is used, correction must be made for the difference in thermal conductivity between nitrogen (N_2) and the gas in use. The gas correction data, charts and tables listed in your controller (such as the Ideal Vacuum Products XGC-320 controller) Operating Manual indicates how different gases affect the display and output from an Ideal Vacuum Products convection gauge.



WARNING! Using a thermal conductivity gauge with gases other than that for which it is calibrated could result in death or serious injury. Be sure to use the correction factor data listed in the controller Operating Manual operating this device.

For N_2 the calibration shows excellent agreement between indicated and true pressure throughout the range from 1×10^{-4} to 1000 Torr. At pressures below 1 Torr, the calibration curves for the different gases are similar. The difference in readings at these low pressures is a constant, a function of the difference between thermal conductivities of the gases.

At pressures above 1 Torr, indicated pressure readings may diverge significantly. At these higher pressures convection currents in the gauge become the predominant cause of heat loss from the sensor and calibration depends on gauge tube geometry and mounting position as well as gas properties.

Generally, air and N₂ are considered the same with respect to thermal conductivity, but even N₂ and air will exhibit slight differences in readings at higher pressures. For example, when venting a system to atmosphere using N₂, you may see readings change by 40 to 55 Torr after the chamber is opened and air gradually displaces the N₂ in the gauge. For most other gases the effect is much more significant and may result in a hazardous condition as described below.

Other considerations when using gases other than N₂ / air

Flammable or explosive gases



WARNING! Ideal Vacuum Products convection gauges are neither intrinsically safe nor explosion proof and are not intended for use in the presence of flammable or explosive gases or vapors.

Under normal conditions the voltages and currents in Ideal Vacuum Products convection gauges are too low to cause ignition of flammable gases. However, under certain failure conditions, sufficient energy could be generated to cause flammable vapors or gases to ignite or explode. Thermal conductivity gauges like the Ideal Vacuum Products convection gauges are not recommended for use with flammable or explosive gases.


Moisture / water vapor

In some processes (lyophilization, for example) the gas composition may not change significantly, except for moisture content. Water vapor can significantly change the response of a thermal gauge and correction should be made, as you would for any other gas.

Other contaminants

If your gases condense, coat, or corrode the sensor, the gauge calibration and response to different gases will change. Generally, if the gauge can be "calibrated" ("zero" and "span" settings), these changes are small enough to be ignored. If you can't set zero and span, the gauge should be replaced or return to factory for evaluation and possible cleaning.

Convection Vacuum Gauge Indicated vs. true total pressure (test gases N₂, Ar, He)



Gas Correction Chart

The Y- axis of the above chart is actual pressure as measured by a capacitance manometer, a diaphragm gauge that measures true total pressure independent of gas composition. The X-axis is the pressure reading indicated by the convection gauge under test. This chart shows readings for an Ideal Vacuum Products convection gauge (CVG) and Granville-Phillips® Convector® gauge to illustrate that the difference in the response for both of these types of gauges is virtually indistinguishable.

CAUTION! Do not assume this data applies to other convection gauges which may or may not be the same. Refer to the correction factor data listed in the controller Operating Manual operating this device. The Ideal Vacuum Products XGC-320 controller Operating Manuals provides a more complete discussion of using correction factors when using the gauge on gases other than Nitrogen.

Ex A: If the gas is nitrogen (N₂), when the true total pressure is 500 Torr, the gauge will read 500 Torr.

Ex B: If the gas is argon (Ar), when the true pressure is 100 Torr, the gauge will read about 9 Torr. If you are backfilling your vacuum system with Ar, when your system reaches a pressure of 760 Torr true pressure your gauge will be reading about 23 Torr. Continuing to backfill your system, attempting to increase the reading up to 760 Torr, you will over pressurize your chamber which may present a hazard.

Ex C: If the gas is helium (He), the controller will read over pressure (OP) when pressure reaches about 10 Torr true pressure and opening the chamber to atmosphere prematurely may present other hazards for both people and product.

CAUTION! What these examples illustrate is that using gases other than nitrogen (N₂) without using accurate gas conversion data and other proper precautions could result in injury to personnel and/or damage to equipment.

Suggested precautions when using gases other than nitrogen (N₂):

Install a pressure relief valve or burst disk on your chamber, to protect it from overpressure. Post a warning label on your gauge readout that states "Do Not Exceed _____ Torr Indicated Pressure" (fill in the blank for maximum indicated pressure for the gas you use) so that an operator using the gauge will not exceed a safe pressure.

5 Service

5.1 Calibration

Every Ideal Vacuum Products XGS-321 gauge is calibrated prior to shipment using nitrogen (N₂). However, you can calibrate the instrument by adjusting zero and span (atmosphere), using the controller which is operating the gauge.

Zero and span (atmosphere) calibration affect the displayed value and the output signal. Zero calibration optimizes performance of the gauge when operating at a low pressure range of 1.00×10^{-4} Torr to 1.00×10^{-3} Torr. If your minimum operating pressure is higher than 1.00×10^{-3} Torr, it is not normally necessary to perform calibration at zero and thus span calibration should be adequate. If you are able to evacuate your system to below 1.00×10^{-4} Torr, it is always a good practice to check and set zero if necessary. This will also improve performance in cases where gauge contamination is causing higher readings than 1.00×10^{-4} Torr, even though the system has been evacuated to below 1.00×10^{-4} Torr. Care should be exercised when using gases other than nitrogen (N₂) / air.

5.2 Maintenance

In general, maintenance is not required for your Ideal Vacuum Products gauge. Periodic performance checks may be done by comparing the gauge to a known reference standard.

5.3 Troubleshooting

Indication	Possible Cause	Possible Solution
Readings appear very different from expected pressure	The process gas is different from the gas used to calibrate the XGS-321	Correct readings for different gas thermal conductivity. See controller Operating Manual
	Module has not been calibrated or has been calibrated incorrectly	Check that zero and span are adjusted correctly. See controller Operating Manual
Readings are noisy or erratic	Loose cables or connections	Check and tighten connections
	Contamination	Inspect gauge for signs of contamination such as particles, deposits, discoloration on gauge inlet. Return to factory for possible cleaning
Gauge cannot be calibrated - zero and span can't be adjusted	Vibration	Ensure gauge is not mounted where excessive vibration is present
		Return to factory for possible cleaning
	Sensor failure for other cause	Return to factory for evaluation or replace
(continued)		

Table "Troubleshooting" (concluded)

Indication	Possible Cause	Possible Solution
Controller displayed pressure is too high and cannot be set to correct value	Contamination	Return to factory for possible cleaning
	Sensor wire damaged	Return to factory for evaluation or replace
Controller displayed pressure is too low and cannot be set to correct value	Sensor wire damaged	Return to factory for evaluation or replace
	Contamination	Return to factory for possible cleaning

5.4 Contamination

The most common cause of all vacuum gauge failures is contamination of the sensor. Noisy or erratic readings, the inability to set zero or atmosphere and total gauge failure, are all possible indications of gauge contamination.

Contamination can be generally characterized as either:

- A) a reaction of process gases with sensor elements, or
- B) an accumulation of material on the sensor elements. Sensors that fail due to chemical reaction are generally not salvageable. Sensors that fail due to condensation, coatings, or particles may possibly be restored by cleaning.

A) Reactive Gases

If process gases react with the materials of construction of the sensor, the result is corrosion and disintegration of the sensor over time. The chemistry of the gases used for plasma etching and other reactive semiconductor processes are examples where this failure mode is possible. In this case, cleaning can't solve the problem because the sensor has been destroyed. The sensor or module must be replaced.

If you experience this failure mode quickly or frequently, you should consider a different vacuum gauge for your application. Thermal vacuum gauges may be available with different sensor materials that are not as reactive with your particular process gases. The standard gold plated tungsten sensor used in the Ideal Vacuum Products convection gauge is offered for use with air and inert gases such as N₂, argon, etc. Ideal Vacuum Products also offers modules with platinum sensors for applications not compatible with gold plated tungsten.

There is no material that is universally chemical resistant; your choice of vacuum gauge (as well as all other vacuum components) should take into consideration the potential reactions between your process gases and the materials of construction. Consider what effect water vapor will have when combined with your process gases because a finite amount of water will enter the chamber during venting to atmosphere with air.

B) Oil, Condensation, Coatings, and Particles

If the failure is due to an accumulation of material in the gauge, we may be able to restore your gauge or module by cleaning. Contamination may be as simple as condensed water, or as difficult as solid particles.

Oils and hydrocarbons: Exposure of the gauge internal surfaces to oils and hydrocarbons can result in sensor contamination. Some of these types of contamination may be removed by cleaning the gauge. If there is the possibility of oil back streaming from wet vacuum pumps, it is recommended that a filter or trap be installed to prevent contamination of components of your vacuum system.

Condensation: Some gases (such as water vapor) can condense on sensor surfaces, forming a liquid coating that changes the rate at which heat is removed from the sensor (which changes the calibration). The sensor can often be restored simply by pumping on the gauge between process cycles. A dry N₂ purge will help speed up drying, or the gauge may be gently heated provided temperature doesn't exceed the specified limit of 40 °C, operating.

Coatings: Some gases can condense on sensor surfaces, forming a solid coating, which changes the rate at which heat is removed from the sensor. Some of these coatings may be removed by cleaning the gauge.

Particles: Particles generated by the process may enter the gauge during the process cycle or during the venting cycle. The result is interference with heat removal from the sensor. In this case, cleaning may be able to remove particles from the gauge. However, particulate contamination is the most difficult to remove as particles can become stubbornly trapped inside the gauge. In some processes, solid particles are created during the process throughout the chamber including inside the gauge. Particles tend to form on cooler surfaces such as in a gauge at room temperature. You may slow down the build-up of particles in the gauge by keeping the gauge warm (within specified limits) during the process cycle.

Particles in the process chamber may be swept into the gauge during the vent cycle. The XGS-321 has a screen built into the gauge port to help keep the largest particles out of the gauge. In very dirty applications, or where particles are small enough to get through the screen, an additional filter installed on the inlet may help prolong the gauge life.

In some vacuum processes, desorbed and sputtered materials from the process may enter vacuum components connected to the process vacuum chamber by line-of-sight transport, especially under high vacuum conditions, i.e., in the molecular flow regime. To prevent materials that may be transported via line-of-sight momentum from entering your vacuum gauge or other components, it is advisable to install some form of apparatus that will block the line-of-sight. In many cases a simple 90° elbow may help prevent or reduce the transport of particles from entering your vacuum gauge.

In the event of gauge contamination please contact the factory to return the gauge for possible cleaning if the gauge has not been exposed to hazardous materials.


5.5 Gauge replacement


If the XGS-321 gauge fails for any reason, and cleaning does not resolve the issue, the XGS-321 gauge should be replaced.

6 Factory Service and Support

If you need help setting up, operating, troubleshooting, or obtaining a return materials authorization number (RMA number) to return the module for diagnosis, please contact us during normal business hours (8:00 am to 5:00 pm Mountain time) Monday through Friday, at (505) 872-0037. Or e-mail us at info@idealvac.com.

7 Returning the Product

**WARNING**



Forwarding contaminated products

Contaminated products (e.g. radioactive, toxic, caustic or microbiological hazard) can be detrimental to health and environment.

Products returned to Ideal Vacuum Products should preferably be free of harmful substances. Adhere to the forwarding regulations of all involved countries and forwarding companies and enclose a duly completed declaration of contamination ^{*)}.

^{*)} Please contact Ideal Vacuum Products

Products that are not clearly declared as "free of harmful substances" are decontaminated at the expense of the customer. Products not accompanied by a duly completed declaration of contamination are returned to the sender at his own expense.

8 Disposal



DANGER

Contaminated parts

Contaminated parts can be detrimental to health and environment.

Before beginning to work, find out whether any parts are contaminated. Adhere to the relevant regulations and take the necessary precautions when handling contaminated parts.



WARNING

Substances detrimental to the environment

Products or parts thereof (mechanical and electric components, operating fluids etc.) can be detrimental to the environment.

Dispose of such substances in accordance with the relevant local regulations.

Separating the components

After disassembling the product, separate its components according to the following criteria:

- Contaminated components
Contaminated components (radioactive, toxic, caustic or biological hazard etc.) must be decontaminated in accordance with the relevant national regulations, separated according to their materials, and disposed of.
- Other components
Such components must be separated according to their materials and recycled.

EU Declaration of Conformity



We, Ideal Vacuum Products, hereby declare that the equipment mentioned below complies with the provisions of the following Directives:

- 2014/30/EU, OJ L 96/79, 29.3.2014
(EMC Directive; Directive relating to electromagnetic compatibility)
- 2011/65/EU, OJ L 174/88, 1.7.2011
(RoHS Directive; Directive on the restriction of the use of certain hazardous substances in electrical and electronic equipment)

Pirani Gauge Enhanced

XGS-321

(operation with XGC-320)

Standards

Harmonized and international / national standards and specifications:

- EN 61000-6-2:2005 (EMC: generic immunity standard)
- EN 61000-6-4:2007 + A1:2011 (EMC: generic emission standard)
- EN 61010-1:2010 (Safety requirements for electrical equipment for measurement, control and laboratory use)
- EN 61326-1:2013; Group 1, Class A (EMC requirements for electrical equipment for measurement, control and laboratory use)

Manufacturer / Signatures

Ideal Vacuum Products, LLC, 87109 New Mexico



Tony C. Smith
General manager

Original: English



XGS-321-100119



New Mexico 87109

Phone: (505) 872-0037

Fax: (505) 872-9001

info@idealtvac.com

www.idealtvac.com